

Caiet de sarcini ATI - Anestezie

Denumire: Circuit ventilatiomecanicaanesteziepentruadulti

Necesar lunar: 1200 bucati

SPITALUL CLINIC JUDELENI DE NEGRUZI
„PIUS BRINZEI”
INTRARE
DATA: 27 MAR. 2023

| Necesar lunar – Anestezie SCJUPB TM | | | | | | |
|-------------------------------------|--|-------|---------------------|--------------------------------|-----------------------|---|
| Nr. crt | Denumire | U. M. | Cantitatenecees ara | Pretestimat iv/ buc (fara TVA) | Pret totaltestimat iv | Specificatiitehnice |
| 1 | Circuit ventilatiomec nica aparat anestezie pentruadulti | buc | 1200 | 45 lei | 54000 lei | <ul style="list-style-type: none"> ✓ Compatibil cu aparatele de anestezie de tip Draeger, General Electric (GE) ✓ Cu certificate/marcajComunitateaEuropenna (CE) ✓ Circuit transparent, steni de unica folosinta. ✓ Ambalatiindividualiin folie ✓ Tubulaturadublade minim 160 cm lungimefiicare brat, cu conectori de 22mm ✓ Balon de ventilatie cu vol.un de 2litri ✓ Tub de minim 80 cm. cu conectori de 22mm pentruatareabalonului de ventilatie ✓ Conector cu cot de 22mm si port de adaptarela gize anestezice (CO2) ✓ Conectoridrepti, in Y ✓ Suprafata interna care saperrnitaeliminaaerulsupara a condensuluiapei in timpul anesteziei ✓ Sa mentinafluxulventilatchiardacatubulaturatesecudat a/ocluzionata accidental |

APROBAT MANAGER
PROF. UNIV. DR. SANDER DOREL



Medic coordonator
Conf.dr. Papurica Marius

Dr. Marian Papurica
Medic primar
anestezie terapie intensiva
Cod 068472

Medic sefsectie ATI
Conf. dr. Bedreag Ovidiu

